GLOVE COMPARISON CHART

SafeTouch™

LATEX GLOVES

	Item	Color	Thickness	Textured	Tensile Strength (After Aging)	Meets Standards	Length	Width	Sterile	Chemo & Fentanyl Tested	Low Protein	Non-Chlorinated	Premium Chlorinated	Polymer Lined	NFPA Certified
Examination Gloves	**2336-2339**		3 mil min.	No	14 MPa, min.	ASTM D3578; ASTM D5151	220 mm minimum for S; 230 mm minimum for M, L, XL	80±10 for S; 95±10 for M; 111±10 for L; 121±10 for XL			●			●	
Sterile Surgical Gloves	**2360-2390**		4 mil min.	Fingertip	18 MPa, min.	ASTM D3577; ASTM D5151	265 mm minimum for all sizes	76±6 for Size 6; 83±6 for Size 6.5; 89±6 for Size 7; 95±6 for Size 7.5; 102±6 for Size 8; 108±6 for Size 8.5; 114±6 for Size 9	●						
Sterile Examination Gloves	**2451-2453**		3 mil min.	Fingertip	14 MPa, min.	ASTM D3578; ASTM D5151	83 mm minimum for S, 95 mm minimum for M, 114 mm minimum for L	80±5 for S; 95±5 for M; 111±5 for L	●		●				
Black Arrow™ Examination Gloves	**2321-2324**		3 mil min.	No	14 MPa, min.	ASTM D3578	220 mm minimum for XS, S; 230 mm minimum for M, L, XL	80 for S; 95 for M; 111 for L; 121 for XL							
Accutouch® Examination Gloves	**6621-6625**		3 mil min.	Fingertip	14 MPa, min.	ASTM D3578	220 mm minimum for XS, S; 230 mm minimum for M, L, XL	70±10 for XS; 80±10 for S; 95±10 for M; 111±10 for L; 121 ±10 for XL			●			●	
Aloetex Examination Gloves	**6716-6719**		3 mil min.	Palm	14 MPa, min.	ASTM D3578	220 mm minimum for XS, S; 230 mm minimum for M, L, XL	70±5 for XS; 80±5 for S; 95±5 for M; 111±5 for L			●				

- All our gloves are FDA 21 CFR Part 820 compliant, and registered with the FDA in their Medical Device Listing database.
- Latex gloves are not recommended for use in the food industry, as some people may be allergic to latex material. Therefore, vinyl gloves are used for most food industry applications.
- Dynarex reserves the right to make changes to these statements

GLOVE COMPARISON CHART

SafeTouch™

NITRILE GLOVES

	Item	Color	Thickness	Textured	Tensile Strength (After Aging)	Meets Standards	Length	Width	Sterile	Chemo & Fentanyl Tested	Low Protein	Non-Chlorinated	Premium Chlorinated	Polymer Lined	NFPA Certified
Blue Nitrile Examination Gloves	**2511-2514**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D5151; ASTM D6978-05	220 mm minimum for S; 230 mm minimum for M, L, XL	80±10 for S; 95±10 for M; 110±10 for L; 120±10 for XL		●					
Black Nitrile Examination Gloves	**2521-2524**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D5151; ASTM D6978-05	220 mm minimum for S; 230 mm minimum for M, L, XL	80±10 for S; 95±10 for M; 110±10 for L; 120±10 for XL		●					
DynaPlus Examination Gloves	**2515-2519**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D6124; ASTM D6978-05	220 mm minimum for XS, S; 230 mm minimum for M, L, XL	70±10 for XS; 80±10 for S; 95±10 for M; 110±10 for L; 120±10 for XL		●					
Sterile Examination Gloves	**6522-6524**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D5151; ISO 11135	220 mm minimum for S; 230 mm minimum for M, L	80±10 for S; 95±10 for M; 110±10 for L	●						
Sterile Examination Gloves	**6528-6530**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D5151; ISO 11135	220 mm minimum for S; 230 mm minimum for M, L	80±10 for S; 95±10 for M; 110±10 for L	●						
Sterile Surgical Gloves	**6537-6539**		6 mil min.	No	12 MPa, min.	ASTM D6319; ASTM D5151; ISO 11135	300 mm minimum for all sizes	102±6 for 8 108±6 for 8.5; 14±6 for 9	●						
AloeSkin Examination Gloves	**6721-6724**		2 mil min.	Fingertip	14 MPa, min.	ASTM D6319; ASTM D6124	220 mm minimum for XS, S; 230 mm minimum for M, L, XL	70±10 for XS; 80±10 for S; 95±10 for M; 110±10 for L; 120±10 for XL						●	

- All our gloves are FDA 21 CFR Part 820 compliant, and registered with the FDA in their Medical Device Listing database.
- Latex gloves are not recommended for use in the food industry, as some people may be allergic to latex material. Therefore, vinyl gloves are used for most food industry applications.
- Dynarex reserves the right to make changes to these statements

GLOVE COMPARISON CHART

SafeTouch™

	Item	Color	Thickness	Textured	Tensile Strength (After Aging)	Meets Standards	Length	Width	Sterile	Chemo & Fentanyl Tested	Low Protein	Non-Chlorinated	Premium Chlorinated	Polymer Lined	NFPA Certified
NITRILE GLOVES															
Gloves in a bag	**2618 2619**		2 mil min.	No	≥14 MPa, min.	ASTM D6319; ASTM D5151	230 mm minimum	110±10 for L 120±10 for XL							
Gloves in a box	**2626UB-2**		2 mil min.	No	≥14 MPa, min.	ASTM D6319; ASTM D5151	230 mm minimum	120±10 for X L							
VINYL GLOVES															
Examination Gloves	**2611-2614**		3 mil min. at palm; 2 mil min. at finger	No	≥13 MPa, min.	ASTM D5250; ASTM D5151	230 mm minimum for all sizes	85±10 for S; 95±10 for M; 105±10 for L; 115±10 for XL						●	
Multi Care Synthetic Examination Gloves	**6812 6814 6815**		3 mil min. at palm; 2 mil min. at finger	No	≥13 MPa, min.	ASTM D5250; ASTM D5151	230 mm minimum for all sizes	85±10 for S; 95±10 for M; 105±10 for L; 115±10 for XL						●	
Next Generation Stretch Examination Gloves	**6824**		3 mil min. at palm; 2 mil min. at finger	No	≥13 MPa, min.	ASTM D5250; ASTM D5151	230 mm minimum for all sizes	85±10 for S; 95±10 for M; 105±10 for L; 115±10 for XL						●	
Gloves in a bag	**2616-2617**		3 mil min. at palm; 2 mil min. at finger	No	≥11 Mpa, min.	ASTM D5250; ASTM D5151	230 mm minimum for all sizes	105±10 for L						●	

- All our gloves are FDA 21 CFR Part 820 compliant, and registered with the FDA in their Medical Device Listing database.
- Latex gloves are not recommended for use in the food industry, as some people may be allergic to latex material. Therefore, vinyl gloves are used for most food industry applications.
- Dynarex reserves the right to make changes to these statements